

INTENDED USE

The Temp-Chex thermometer assembly is designed as a convenient device for monitoring the temperature when storing a variety of products including clinical laboratory reagents, controls, and patient specimens.

PRINCIPLE

A glass thermometer is the main component of the system. It is manufactured to meet NIST standards of accuracy and is certified traceable to NIST and ASTM E77 standards. Temp-Chex is accurate to ± 1 °C. Temp-Chex is not a NIST thermometer.

PRECAUTIONS

SDS can be obtained at streck.com, by calling Streck at 800-843-0912, or by calling your local supplier.

COMPONENTS

A short stem bulb immersion thermometer is inserted through a rubber stopper into a 10ml vial containing a solution of glycol. The thermometer and vial are enclosed in a plastic sleeve; **DO NOT REMOVE FROM THE PLASTIC SLEEVE.** The plastic sleeve is a safety feature and will contain fluid and glass should breakage occur. This assembly can be attached magnetically or adhesively to surfaces.

DEFOGGING OF SLEEVE

Occasionally when the Temp-Chex sleeve is exposed to warm air, a fog may form on the sleeve making it difficult to read the thermometer. Streck has treated the sleeve with an anti-fogging material. Over time this material can lose its effectiveness. If the plastic sleeve is fogging, Streck recommends an application of ordinary eye glass defogger to the outside of the sleeve.

CALIBRATION STATEMENT

Prior to shipment, Temp-Chex thermometers are stored in a controlled environment. With proper care, the thermometer should maintain its accuracy. Due to use in diverse applications and handling, there is no precise way to predict how long calibration will be maintained. As with any thermometer, drift is a possibility and accuracy can be affected by shock, aging, temperature and contamination through daily use. **Periodic recalibration or verification should be scheduled by each laboratory in accordance with its accrediting agency and/or established procedure.** At a minimum, annual recalibration or verification from the date the thermometer is put "in use" is good lab practice. It is recommended to record the "in use" date on the Calibration Certificate for documentation purposes. The Calibration date listed on the Calibration Certificate for this thermometer is for reference use only. Please contact Streck Technical Services at 800-843-0912 or technicalservices@streck.com with any questions.

INSTRUCTIONS FOR USE

1. **Visually inspect the thermometer immediately after unpacking.** The Temp-Chex can be placed directly into service provided there is no fluid separation inside the thermometer column. Gaps or breaks in the fluid are not a product defect. **If the fluid has separated, see Directions for Uniting Fluid Separation.**
2. Mounting Instructions
 - a. Temp-Chex must be stored in the upright position, bulb down. Prior to attachment, determine if the wall is ferrous metal and will allow use of the magnetic section. If so, make sure the surface is clean and free of any moisture or frost. To utilize the magnet, remove the second magnet from the magnet already affixed to the Temp-Chex and attach the entire Temp-Chex to the surface. Verify that it will hold properly.
 - b. If the wall is not ferrous metal or if the surface does not provide proper magnetic attraction, the adhesive surface of the magnets can be used. Once again, the surface should be clean and dry. First check that the two magnetic parts adhere to each other (if incorrectly aligned, the magnets will repel each other). Maintain proper orientation of the magnetic strip when attaching it to the freezer, incubator or refrigerator wall. Remove paper backing from the second magnet, attach the adhesive side to the wall and allow a few minutes to establish the appropriate bonding prior to attaching the magnet already affixed to the Temp-Chex. It is safest to attach the Temp-Chex to walls in such a position that the Temp-Chex can also be supported by a shelf. However, provided the appropriate attachment procedure has been followed, the Temp-Chex should remain stationary.

DIRECTIONS FOR UNITING FLUID SEPARATION

Occasionally, fluid separation occurs during shipping or during normal use. This separation is evidenced by single or multiple breaks in the column, fluid reservoir at the bottom or expansion bulb at the top of the thermometer. **If any of these are obvious or if the thermometer does not seem to be functioning properly, the following procedures should be followed to reunite the fluid.**

Before attempting the following methods, the thermometer must be removed from the protective sleeve and glycol-filled vial. To remove the thermometer from the vial, hold the vial in an upright position and use a gentle twisting motion. CAUTION: THE THERMOMETER IS MADE OF GLASS AND GLASS CAN BREAK. DONNING OF SAFETY GLASSES AND GLOVES IS RECOMMENDED. EITHER WRAP THE THERMOMETER IN PARAFILM[®] OR WEAR RUBBER GLOVES BEFORE REMOVING THE THERMOMETER FROM THE VIAL.

HEATING METHOD

Heat the thermometer bulb in an upright position away from your face in **warm liquid, air or over a soft flame.** Allow the liquid column to rise slowly until the separated position of the column enters the expansion chamber at the top of the thermometer. **NOTE: FILLING THE EXPANSION CHAMBER MORE THAN HALFWAY WILL BREAK THE THERMOMETER.** Tap the thermometer GENTLY (if struck too hard,

the fluid reservoir can break) while allowing the gas separating the column to rise to the top of the chamber. Allow the thermometer to cool slowly in an upright position.

After using the above method, examine the thermometer for fluid separation. **Repeat the procedure if separation is still evident.** When the fluid is united, carefully reinsert the thermometer into the glycol-filled vial and replace the unit in the protective sleeve. It is good laboratory practice to check the thermometer against a thermometer of known accuracy after the procedure to unite the fluid is complete.

If the fluid cannot be reunited, contact Streck Technical Services at 800-843-0912. Be prepared to provide the catalog number and serial number listed on the Calibration Certificate.

LOW TEMPERATURE THERMOMETERS

According to the U.S. Department of Commerce monograph covering liquid-in-glass thermometry, the use of mercury-in-glass thermometers for low temperature measurements is limited by the freezing point of mercury. Organic liquids, such as red spirit, may have errors in temperature measurement if sufficient precautions are not taken. Organic liquids may wet the glass surface and leave a film on the wall of the capillary when the liquid column recedes. The thickness of the film on the capillary wall will depend, among other things, on the viscosity of the liquid, the interfacial action between the liquid and glass, and the rate at which the thermometer is cooled. Whenever possible, the rate of cooling should be slow with the bulb cooled first, enabling the viscosity of the organic fluid in the capillary to be kept as low as possible until the final temperature is reached. This minimizes the amount of liquid left on the capillary wall. Sufficient time should always be allowed to ensure complete drainage. This may take an hour or more under adverse conditions.

ORDERING INFORMATION

Please call our Customer Service Department at 800-228-6090 for assistance. Additional information can be found online at streck.com.

All product names, logos, brands, and marks are property of their respective owners.